

Underwriters Laboratories (UL LLC) Safety Certification Report



Model: TR100Mx, where x equals 120, 150, 180, 190, 200, 240 or 480 to represent different output ratings.

Device Description: Switching Adapter

Applicant: CINCON ELECTRONICS CO LTD
8-1 FU KUNG RD, FU HSING PARK, FU HSING HSIANG,
CHANGHUA HSIEN, 506, TAIWAN

Manufacturer: Same as Applicant

Manufacturing Facility(ies): DONGGUAN CINCON ELECTRONICS LTD
66 Yuanying Rd
Liaobu Town
Dongguan
Guangdong 523412 CHINA

CINCON ELECTRONICS CO LTD
8-1 FU KUNG RD, FU HSING PARK, FU HSING HSIANG,
CHANGHUA HSIEN, 506, TAIWAN

Report No.: E252331-D1006-2/A0/C0-UL

Report (Re)Issue Date: 2024-01-03

Base Standard(s): ANSI/AAMI ES60601-1:2005/(R)2012 and A1:2012, C1:2009/(R)2012 and A2:2010/(R)2012, CAN/CSA C22.2 No. 60601-1:14

Additional Standards: • AAMI ES60601-1:2005/(R)2012 and A1:2012/(R)2012 and A2:2021;
• CAN/CSA-C22.2 No. 60601-1:14 (Reaffirmed 2022)

Report Types: This report consists of the following report types:
- [Yes] US Certification (UL Recognition)
- [Yes] CAN Certification (cUL Recognition)

This report covers the Safety evaluation of the referenced model(s) according to the standard(s) specified above.